### 510(k) Summary

510(k) Summary as required by <u>21 CFR § 807.92</u>

510(k) Submitter:	Cardinal Health 207, Inc. Yorba Linda, CA 92887 Phone: (714) 283-8472 Fax: (714) 283-8472				
Contact Person	Andre von Muller phone/fax: (714) 292-9464 email: andre.vonmuller@cardinalhealth.com				
Establishment Registration Number	2050001				
Date prepared	June 4, 2008				
Name of the device	THE ADVANTAGE SERIES® Non-vented (NV) Full Face Mask				
Common/usual name	Mask for use with ventilator (continuous, facility use)				
Classification	THE ADVANTAGE SERIES® Non-vented (NV) Full Face Mask is classified as a Class II device under the following classification code:				
	Product Code	CFR Section	Panel		
	СВК	21 CFR <u>868.5895</u>	Anesthesiology		
Reason for the submission			Cardinal Health 207, Inc.	· · · · · · · · · · · · · · · · · · ·	
Substantially equivalent device	THE ADVANTAGE SERIES® Non-vented (NV) Full Face Mask is substantially equivalent to the following device:  > Image3 SE Disposable Full Face Mask (K023135).				
	similarities to the	predicate device: ended use	ed (NV) Full Face has the foll	owing	
	Labeling for the p	oredicate device to whe	ich substantial equivalence has 510(k) submission.	as been	
Device Description.	seal around the r source is directed adjustable head connected to a r 360°. The elbow positive pressure (EN1281-1) with This mask can be requiring pressure	nose and mouth such d into the patient's nos gear. The design cons nolded polycarbonate shall have an intercone device. The intercone a 22mm female conne used to deliver posities up from 3 to 40 cm	ted (NV) Full Face Mask provi that pressure from a positive page and mouth. It is held in place ists of latex-free silicone cush frame with an elbow that can inection mating with the tubing nection shall conform to ISO 5 ector that can also swivel thro live pressure therapy to patien in H <sub>2</sub> O. This mask can be used el of pressure and that have a	pressure be with an ion swivel g from the 5356-1 ugh 360°. hts with a	

mechanism to adequately remove exhaled gases as well as a safety valve that opens to atmosphere to provide room air in the event of loss of supply pressure. THE ADVANTAGE SERIES® Non-vented (NV) Full Face Mask is functionally similar to Cardinal Health Advantage Series Full Face mask (K043382) except that it does not have the anti-asphyxia or an exhalation port. Due to the fact that this is a non-vented mask, there is a need for a separate mechanism to remove exhaled gases. Reference Section 22 of this submission for a drawing of the mask.

### Intended Use/Indications for

THE ADVANTAGE SERIES® Non-vented (NV) Full Face Mask is intended to provide a patient interface for application of noninvasive ventilation. The mask is to be used as an accessory to ventilators which have adequate alarms and safety systems for ventilator failure, and which are intended to administer CPAP or positive pressure ventilation for treatment of respiratory failure, respiratory insufficiency, or obstructive sleep apnea.

The mask is disposable and for single patient use. It is for use on adult patients (> 30 kg), who are appropriate candidates for noninvasive ventilation and use in a hospital/institutional environment.

#### Comparison of technological characteristics between devices.

THE ADVANTAGE SERIES® Non-vented (NV) Full Face Mask is functionally similar to the Cardinal Health Advantage Series Full Face mask (K043382) except that it does not have the anti-asphyxia or an exhalation port. It is constructed from the same materials as the Advantage Series Full Face Mask.

THE ADVANTAGE SERIES® Non-vented (NV) Full Face Mask is substantially equivalent to the Image3 SE Disposable Full Face Mask (K023135). A comparison of the Advantage Full-Face Non-Vented (NV) Mask to the Image3 SE Disposable Full Face Mask has been provided in Section 12 of this 510(k) submission.

### Summary of non-clinical performance testing.

THE ADVANTAGE SERIES® Non-vented (NV) Full Face Mask was tested under various conditions using the Cardinal Health design control process to evaluate design parameters. Testing included:

- Conformance to design specifications
- Interconnection to patient circuit compliant with international standards
- Environmental testing
- **Dead Space**
- Pressure vs. Flow
- Leak Allowance
- Cleaning

The mask passed the specified test criteria.

## performance experience.

Summary of clinical THE ADVANTAGE SERIES® Non-vented (NV) Full Face Mask was not subjected to human clinical studies to validate the performance of the device.

#### Conclusion

The bench performance data of the ADVANTAGE SERIES® Non-vented (NV) Full Face Mask, when compared to the data and/or claims made on the predicate devices, demonstrate that the device is as safe, as effective, and performs as well as or better than the predicate devices. The intended use of the ADVANTAGE SERIES® Non-vented (NV) Full Face Mask is the same as the predicate device. No new questions of safety or effectiveness are raised.



Food and Drug Administration 9200 Corporate Boulevard Rockville MD 20850

OCT 09 2008

Mr. Andre von Muller Senior Regulatory Affairs Engineer Cardinal Health 207, Incorporated 22745 Savi Ranch Parkway Yorba Linda, California 92887

Re: K081670

Trade/Device Name: ADVANTAGE SERIES® Non-vented (NV) Full Face Mask

Regulation Number: 21 CFR 868.5895 Regulation Name: Continuous Ventilator

Regulatory Class: II Product Code: CBK

Dated: September 26, 2008 Received: September 29, 2008

Dear Mr. von Muller:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the <u>Federal Register</u>.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050. This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Center for Devices and Radiological Health's (CDRH's) Office of Compliance at (240) 276-0120. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding postmarket surveillance, please contact CDRH's Office of Surveillance and Biometric's (OSB's) Division of Postmarket Surveillance at 240-276-3474. For questions regarding the reporting of device adverse events (Medical Device Reporting (MDR)), please contact the Division of Surveillance Systems at 240-276-3464. You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <a href="http://www.fda.gov/cdrh/industry/support/index.html">http://www.fda.gov/cdrh/industry/support/index.html</a>.

Sincerely yours,

Chiu Lin, Ph.D.

Director

Division of Anesthesiology, General Hospital, Infection Control and Dental Devices

Jechmuels Lud mis for11

Office of Device Evaluation

Center for Devices and

Radiological Health

# Indications for Use

nown): K081670		
ADVANTAGE SERI	ES® Non-vented	(NV) Full Face Mask
provide a patient inte to be used as an acc systems for ventilatio positive pressure ven insufficiency, or obstr The mask is disposal (>30 kg) who are app	rface for applicati essory to ventilate in failure and which itilation for treatm ructive sleep apno- ble and for single propriate candidate	
lbpart D) VRITE BELOW THIS L	AND/OR  INE- CONTINUE	Over the Counter Use(21CFR 801 Subpart C)
(Division Sign- Division of Ane	Off) esthesiology, Gene	ral Hospital
	e: The ADVANTAGE is provide a patient interest to be used as an accessive pressure ver insufficiency, or obstitutional (>30 kg) who are appropriately institutional (>30 kg) who are appropriately instit	provide a patient interface for applicat to be used as an accessory to ventilat systems for ventilation failure and which positive pressure ventilation for treatments insufficiency, or obstructive sleep application for single (>30 kg) who are appropriate candidate hospital/institutional environment.  AND/OR